



Ref. Certif. No.

CH-8456

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME**CB TEST CERTIFICATE**

Product

Optical 3D measurement system

Name and address of the applicant

Atracsys Sàrl Route du Verney 20 SWITZERLAND
1070 Puidoux

Name and address of the manufacturer

Atracsys Sàrl Route du Verney 20 SWITZERLAND
1070 Puidoux

Name and address of the factory

Atracsys Sàrl Route du Verney 20 SWITZERLAND
1070 Puidoux*Note: When more than one factory, please report on page 2* Additional Information on page 2

Ratings and principal characteristics

48 VDC, 0.6 A
Class II, IP20

Trade mark (if any)

Fusion Track

Customers's Testing Facility (CTF) Stage used

Model / Type Ref.

fTk500/fusionTrack 500

Additional information (if necessary may also be reported on page 2)

A sample of product was tested and found to be in conformity with IEC

 Additional Information on page 2IEC 60601-1:2005,
IEC 60601-1:2005/AMD1:2012
IEC 60601-1-2:2014
IEC 62304:2006,
IEC 62304:2006/AMD1:2015

National differences

EU Group Differences;
EU Special National Conditions;
EU A-Deviations;
CA; JP; US

As shown in the Test Report Ref. No. which forms part of this Certificate

16-MO-0108.S01 + .S40 + .E01(EMC) + .E02(EMC)

This CB Test Certificate is issued by the National Certification Body

Electrosuisse
Luppenstrasse 1
8320 Fehraltorf
SWITZERLANDSigned by: Martin Plüss
Date: 2017-01-04

page 1 of 1