

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Optical 3D measurement system

Name and address of the applicant

Atracsys Sàrl
Route du Verney 20, 1070 Puidoux
Switzerland

Name and address of the manufacturer

Atracsys Sàrl
Route du Verney 20, 1070 Puidoux
Switzerland

Name and address of the factory

Atracsys Sàrl
Route du Verney 20, 1070 Puidoux
Switzerland

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

48 VDC, 0.6 A (POE+)
Class II, IP20

Trademark (if any)

atracsys

Customer Test Facility (CTF) Stage used

./.

Model / Type Ref.

fusionTrack 250 / fTk250
fusionTrack 500 / fTk500

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012
IEC 60601-1-2:2014

National differences:

EU Group Differences
EU Special National Conditions
EU A-Deviations

As shown in the Test Report Ref. No. which forms part of this Certificate

19-MO-0017.E01, .E02, .E03, .E04, .S01

This CB Test Certificate is issued by the National Certification Body



Electrosuisse Product Testing AG

Luppenstrasse 3

8320 Fehraltorf

SWITZERLAND

Date: 2019-04-16

Signature: Martin Plüss